



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R60/1991-GB-94.08

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue**
Teddington
Middlesex
TW11 0JZ

Person responsible: **M D Koch - Head of Type Examination**

Applicant

Name: **Railweight.**
Address: **Hurstfield Industrial Estate.**
Hurst Street
Reddish
Stockport
SK5 7BB

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

Load Cell Model No Weighline RWT -X
Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60 edition 1991 (E) for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report no TR00214, that includes 35 pages.

The issuing authority

NWML

The OIML member



S J Bennett

Date: 12 August 1994

Date: 12 August 1994

Characteristics : Accuracy class C

Capacity E_{\max}	v_{\min}	E_{\min}	n_{\max}
1500 kg	0.333 kg	150 kg	750
5000 kg	1.111 kg	500 kg	750
7500 kg	1.667 kg	750 kg	750
10000 kg	2.222 kg	1000 kg	750
12500 kg	2.778 kg	1250 kg	750
15000 kg	3.333 kg	1500 kg	750

Important note: Apart from the mention of the certificate reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.